

FIG. 2




Fig. 3

PCA APPLICATION	INTERMITTENT APPLICATION	TPN APPLICATION	CONTINUOUS APPLICATION			
PCA CODE	INTERMITTENT CODE	TPN CODE	CONTINUOUS CODE	BOOT SYSTEM	TERMINAL UTILITY	TEST UTILITY
PUMP APPLICATION TEMPLATE	PUMP APPLICATION TEMPLATE	PUMP APPLICATION TEMPLATE	PUMP APPLICATION TEMPLATE	BOOT	TERMINAL	TEST
PUMP KERNAL	PUMP KERNAL	PUMP KERNAL	PUMP KERNAL	PUMP KERNAL	PUMP KERNAL	PUMP KERNAL
SERIAL COM. PROTOCOL	SERIAL COM. PROTOCOL	SERIAL COM. PROTOCOL	SERIAL COM. PROTOCOL	SERIAL COM. PROTOCOL	SERIAL COM. PROTOCOL	SERIAL COM. PROTOCOL
PUMP DRIVERS	PUMP DRIVERS	PUMP DRIVERS	PUMP DRIVERS	PUMP DRIVERS	PUMP DRIVERS	PUMP DRIVERS

400a 400b 400c 400d 400e 400f 400g

402

CORE BANK

FIG. 4

FIG. 4




FIG. 5

09920467.080101

310

App 4 Configuration Parameters
App 3 Configuration Parameters
App 2 Configuration Parameters
App 1 Configuration Parameters
App Template Configuration Parameters
Launch App Delivery/Status Parameters
KERNAL DATA: Automatic Odometers, Program Lock-up Table, Launch Program Select, Kernal and App Need Defaulting flags
DRIVERS DATA: Error Log
DRIVERS DATA: Manufacturing Parameters

FIG. 6




FIG. 7A




FIG. 7B

0920467 0804266




FIG. 7C




FIG. 8

Batch Programming - Default Protocol

File Configure Connection Window Help

Delivery & Options

Delivery

Programming Units: 928
Concentration (mg/ml): 904
Continuous Rate (mg/hr): 906
Demand Dose (mg): 908
Continuous Demand Dose (mg/hr): 926
4 Hour Delivery Limit: 912
☐ No Limit
☒ Delivery Limit (mg): 914

Options

Lock Level: 930
Epidural Mode: ☐ 922
Air Detector On: ☐ 918
Time/Date: 902
☒ Set Time/Date

Reservoir Volume

☒ In Use
☒ Reset to (ml): 916

924

910

90

FIG. 9